

Product Name :
Accessories For Compression Testing Machine Cement
Mortar Flexural Testing

Product Code :
CHINAELABC5450001



Description :

Accessories For Compression Testing Machine Cement Mortar Flexural Testing

Technical Specification :

The flexural Jig is designed for testing of 40 x 40 x 160mm cement prism
The frame is fitted with roller bearings accordance to EN 196-1.
Robust frame fitted with an upper bearer that moves vertically supported by springs.
One of the two lower bearer can tilt horizontally and the distance between the two bearers is 100 mm.
Two lower bearer (supporting rollers) are of Length between 45-50 mm, having $(10 \pm 0,5)$ mm diameter.
The upper bearer also has same length between 45-50 mm, having $(10 \pm 0,5)$ mm diameter.
Total Height 188 mm.